

## IMPORTANT INFORMATION TO READ and RETURN

## Installation Requirements for a Whitley A135 HEPA Anaerobic Workstation

Thank you for choosing one of our products for your laboratory. To enable our engineers to perform an efficient, trouble-free installation please study, complete and return this form to us by fax: 01274 531197 or email: service@dwscientific.co.uk. Should you have any questions, please do not hesitate to contact us, as we are here to help. When we have received the completed form, our Service Department will contact you to arrange a mutually convenient installation date.

The following information represents the ideal requirement.

Please contact us IMMEDIATELY if your intended location does not match this specification.

|                  |                    |  |  | $\checkmark$ |  |  |  |  |
|------------------|--------------------|--|--|--------------|--|--|--|--|
| Access Requ      | uirements          |  |  |              |  |  |  |  |
| Please note that | at this product is | s wider than a   | standard door so please ensure that you consider access carefully. |              |  |  |  |  |
| Space Requi      | rements            |  |  |              |  |  |  |  |
| •                |                    | quipment is 175 kg. This workstation comes complete with a bespoke trolley. However, if you wish to locate the workstation contact us to discuss the implications. |  |              |  |  |  |  |
| External Dim     | ensions*           |  |  |              |  |  |  |  |
| Width            | Depth              | Height   |  |              |  |  |  |  |
| mm               | mm                 | mm   |  |              |  |  |  |  |
| •                |                    |  |  |              |  |  |  |  |

## \* Please Note:

1452

1056

993

If benchmounted, a further minimum clearance of 500mm is required above the unit and a minimum clearance of 500mm on both sides of the unit is required for user/service access.

If the **Refrigeration System** is fitted, please allow a clearance of 770mm behind the workstation. If the **Heat Removal System** is fitted, please allow a clearance of 600mm behind the workstation. Some of this clearance is necessary for air circulation.

## **Gas Requirements**

The incoming gas supplies must be terminated near the right hand side of the main chamber and fitted with leak-proof taps and pressure gauges.

The gas lines to which the equipment is attached are the responsibility of the user and should be constructed, tested and maintained to the standards specified within the British Compressed Gasses Association (BCGA) Code of Practice CP4 (or international equivalent). Gas lines previously used for flammable gases must be purged prior to re-use.

Regulators should be fitted in accordance with the information contained in the table below and the various pressures strictly adhered to.

| Gas Type             | Connection Details   | Cylinder Regulator<br>Required | Regulator Outlet Range | Flow Rate  |
|----------------------|--|--------------------------------|------------------------|--|
| Anaerobic<br>Gas Mix | 1/4" BSP male fitting or connection for 6mm Polyurethane tubing.  Hydrogen/Anaerobic Gas Mixture Regulator – Two Stage – order Code A01745 | Two stage                      | 4 - 6 bar (60-90 psi)  | 40 litres per<br>minute<br>(dynamic)             |
| Nitrogen             | 1/4" BSP male fitting or connection<br>for 8mm Polyurethane tubing<br>Nitrogen Regulator – Two<br>Stage – order Code <b>A01748</b>         | Two Stage                      | 4 - 6 bar (60-90 psi)  | Minimum<br>150 litres per<br>minute<br>(dynamic) |

DWS/SSP0005 Form Revision Level: 04

| Connection Type   |  |  |  |  |  |  |  |  |
|---|--|--|--|--|--|--|--|--|
|   | Push in connection  OR                   |  |  |  |  |  |  |  |
|   | Push on Connection                       |  |  |  |  |  |  |  |
| Mains Requirements  |  |  |  |  |  |  |  |  |
| Electricity Supply  | Wall Socket                              |  |  |  |  |  |  |  |
| 240 volts   | 1 x Three Pin, 13 Amp. Minimum 6A rating |  |  |  |  |  |  |  |
| Other Considerations  Do not site the instrument near draughts caused by windows, doors or air conditioning systems  Remember, if you do not have the required regulators you can order them from Don Whitley Scientific::  • Hydrogen/Anaerobic Gas Mixture Regulator – Two Stage – order Code A01745  • Nitrogen Regulator – Two Stage – order Code A01748  |  |  |  |  |  |  |  |  |
| Notes  If an existing unit is being taken in part exchange or is being removed from the laboratory, it must be de-contaminated before DWS staff handle the unit.  |  |  |  |  |  |  |  |  |
| In the UK, delivery and installation are free of charge (unless otherwise agreed). If our engineers are unable to install the unit and a return journey is necessary, <b>a charge may be made</b> . Export customers, please refer to your local distributor.  It is essential that this form is completed and returned, to avoid delay to your installation. |  |  |  |  |  |  |  |  |
| THANK YOU FOR THINKING WHITLEY  |  |  |  |  |  |  |  |  |
|   |  |  |  |  |  |  |  |  |
|   |  |  |  |  |  |  |  |  |
| Signature   | Title                                    |  |  |  |  |  |  |  |
| Print Name  | Establishment                            |  |  |  |  |  |  |  |